

The Zeiss logo, consisting of the word "ZEISS" in white capital letters on a black rectangular background.

Fax Cover Sheet

Surveying Division

To: John Till
Company: N.G.S.
Phone:
Fax: 301-713-3232

From: Robert J.W. Siemers
Date: August 8, 1997
Phone: 914 681 7303
Fax: 914 681 7472
Pages: 3

Subject: Bar code Calibration

EXAMPLE #4 - Page 1 of 3

Geomatics Canada - NRCan
Geodetic Survey Division
Rod Calibration Report

← GSC

for: CARL ZEISS INC., NEW YORK

December, 1995

Rod Type: NEDO-ZEISS 3 metre invar
Serial Number: 15254
Air Temperature: 24°C

Barcode Rod (Single Scale)

Nominal
(millimetres)

Measured Distance
(millimetres)

125.0	125.105
215.0	215.099
325.0	325.098
425.0	425.095
505.0	505.088
605.0	605.093
715.0	715.095
805.0	805.102
915.0	915.101
1025.0	1025.103
1105.0	1105.104
1185.0	1185.104
1325.0	1325.107
1405.0	1405.109
1505.0	1505.112
1595.0	1595.111
1705.0	1705.116
1805.0	1805.117
1925.0	1925.117
2025.0	2025.119
2145.0	2145.123
2225.0	2225.121
2295.0	2295.123
2395.0	2395.121
2515.0	2515.120
2585.0	2585.121
2705.0	2705.123
2795.0	2795.118
2865.0	2865.119
2965.0	2965.115

EXAMPLE #4 - Page 2 of 3

NEDO-ZEISS BARCODE ROD # 15254

FOR: CARL-ZEISS INC., NEW YORK, U.S.A

